

The Ministry of Education Culture, Research, and Technology (MoECRT)

Strengthen University Teaching Hospitals in the Fight against COVID-19 and Pandemic Preparedness

Supply, Installation, Testing, Commissioning and User Training of Medical and Laboratory Equipment and Furniture for the Brawijaya University Hospital (RSUB), Malang, East Java, Republic of Indonesia

ACA/2021/427-609 / 515000377 / GOODS

Circular 1

N°	Question	Answer
1	<p>I am interested to bid on ICB: ACA/2021/427-609 / 515000377 / GOODS - Procurement of Medical/Laboratory Equipment and Furniture.</p> <p>Can you send me the bidding document by e-mail?</p>	<p>Please refer to the Invitation for Bids: "A complete set of bidding documents can be downloaded by the interested Bidders from the following websites: LPSE website: https://lpse.kemdikbud.go.id/eproc4/ UB Website: https://piuphln.ub.ac.id/Tender/ The Purchaser will publish its responses and clarifications with reference to ITB 7.1. [Clarification of Qualification and Bidding Documents, Site Visit, Pre-Bid Meeting] on the websites mentioned above. All bidders are advised to check this web pages daily to receive all clarifications and responses in a timely manner."</p>
2	<p>Can you please send us the tender file in a workable format (Word, Excel) so we can extract the forms, specifications and other information we need to prepare our Bid?</p>	<p>Yes, the technical specifications, BoQs and the Bidding Document will be uploaded to the website mentioned in the Bidding Document Section II, ITB 7.1, for the Bidders' information only, with the intention to support preparation of Bids.</p> <p>The Purchaser does not take any responsibility for correctness or accurateness of the submitted information.</p> <p>The original Bidding Document prevails in case of contradictions.</p>
3	<p>In section VII, List of Goods and Delivery Schedule, it is not clearly indicated what the requested delivery delay is. Can you please inform us what it is? Now it says 120. Do you mean 120 days? If so, 120 days from what?</p>	<p>Please refer to Section VII, 1. List of Goods and Delivery Schedule, clearly indicating "(...) number of days following the date of signature of the Contract".</p>
4	<p>In the tender file, Section VII. Schedule of Requirements, (3. Technical Specifications) page 79, it is indicated that we need to include in our offer UPS protection for electro-</p>	<p>We understand the Bidder refers to UPS devices, not UPC devices.</p>

N°	Question	Answer
	sensitive items. Please indicate clearly for which items we need to include UPC devices and for which items not.	<p>Please refer to Section VII, 3. Technical Specifications, stating as follows</p> <p><i>“Common Specifications & Requirements</i></p> <p><i>All equipment must be supplied with their standard accessories as normally provided by the manufacturers and with those accessories that are specifically mentioned in the item specifications. Cost of these accessories must be included in the tender price. In the bidding documents, bidders should provide a detailed list of these accessories and recommend those additional accessories that are deemed necessary. All mains electrically powered items must be suitable for operation using the electrical system within Indonesia (220-240 volts 50-60 Hz). Plugs are to be 3-pins fused, up to 16 Amps where applicable. Cable connections must be at least 3 metres long, measured from outside of the equipment. For electro-sensitive items such as the ultrasound or anaesthesia machine or ventilators UPS protection should be included in the offer.”</i></p> <p>It is the Suppliers responsibility to scrutinize Annex 1 Technical Specifications and Annex 3 BoQ to identify the electro-sensitive equipment items and to propose the required UPS devices as part of their Bid.</p>
5	Please send us your Tax Identification Number. We need it for DHL, to send you our BID.	The Tax Identification Number of the Purchaser is not required to send a bid.

